



시험성적서

1. 성적서 번호 : CT19-089653K
2. 의뢰자
 - 업체명 : 창영메디칼주식회사
 - 주소 : 충청북도 음성군 금왕읍 초금로 704-27 (오선리)
3. 시험기간 : 2019년 08월 02일 ~ 2019년 08월 28일
4. 시험성적서의 용도 : 품질관리
5. 시료명 : TPE Tube
6. 시험방법 / 시험장소
 - (1) 의뢰자 제시방법(KS P 1907-1) > 인천광역시 연수구 갯벌로 145번길 8

확인	작성자 성명	최동규		기술책임자 성명	정규진	
비고 : 1. 이 성적서는 KS Q ISO/IEC 17025 및 KOLAS 인정과 관련이 없으며, 의뢰자가 제시한 시료 및 시료명에 한정된 결과로서 전체제품에 대한 품질을 보증하지는 않습니다. 2. 이 성적서는 홍보, 선전, 광고 및 소송용으로 사용될 수 없으며, 용도 이외의 사용을 금합니다. 3. 이 성적서의 일부만을 발췌하여 사용한 결과는 보증할 수 없습니다. 4. 이 성적서의 진위여부는 홈페이지(www.kcl.re.kr)에서 확인 가능합니다.						

2019년 08월 28일




한국건설생활환경시험연구원



결과문의 : 21999 인천광역시 연수구 갯벌로 145번길 8 ☎ (032)713-5250

총 27페이지 중 1페이지

양식TQP-12-01-03(0)



시험성적서

성적서번호 : CT19-089635K

7. 시험결과

1) 시험방법(KS P 1907-1)

(1) 표준물질 및 시험물질

- ① Nitroglycerin(순도: 0.974 %, Lot No. H0J138, USP, USA)
- ② Cyclosporin A(순도: 99.9 %, Lot No. C574E, TCI, Japan)
- ③ Tacrolimus(순도: 95.1 %, Lot No. LRAC1117, Sigma-Aldrich, USA)
- ④ 페링가니트 0.1% 주사(Nitroglycerin 1.0 mg/mL, Lot No. 5771202, Elyson, Germany)

(2) 표준물질 검량선 조제농도

- ① Nitroglycerin: 1.000 µg/mL, 5.000 µg/mL, 10.000 µg/mL, 20.000 µg/mL, 100.000 µg/mL
- ② Cyclosporin A: 5.000 µg/mL, 10.000 µg/mL, 50.000 µg/mL, 100.000 µg/mL, 200.000 µg/mL,
1000.000 µg/mL
- ③ Tacrolimus: 2.500 µg/mL, 5.000 µg/mL, 10.000 µg/mL, 15.000 µg/mL, 20.000 µg/mL

(3) 시험군의 구성

시험물질	시험물질 조제농도(µg/mL)	시험용액 채취 시간(h)
Nitroglycerin	100.000	
Cyclosporin A	50.000	1 h, 2 h, 3 h, 4 h
Tacrolimus	10.000	

(4) 시험방법: 5 % 넥스트로스용액을 사용하여 조제된 시험물질(Nitroglycerin, Cyclosporin A, Tacrolimus)을 정량이송펌프를 이용하여 시험물질(약물)에 따라 주입 속도를 설정한다. 약물 수액이 수액 튜브를 통과하여 처음 회수되기 시작하는 시점을 0으로 하고 각 시점(1 h, 2 h, 3 h, 4 h) 별 시험용액 10 mL를 회수한다. 회수된 시험용액은 0.2 µm syringe filter로 여과 후, 메탄올로 회석 전처리하여 HPLC 분석하여 약물을 측정한다.

----- 다음페이지 계속 -----

시험성적서

성적서번호 : CT19-089635K

(5) 약물흡착도 산출식

$$\text{약물흡착도(%)} = \frac{C_0 - C_p}{C_0} \times 100$$

C_0 : 초기 조제한 수액의 약물 측정농도($\mu\text{g/mL}$)

C_p : 수액 투브 통과 후 수액의 약물 측정농도($\mu\text{g/mL}$)

2) 기기분석방법

(1) 기기 및 시약

- ① HPLC: YL9100(Younglin, Korea)
- ② 저울: EPG 214C(OHAUS, USA)
- ③ 정량이송펌프: Model7518-00, Masterflex, USA
- ④ Syringe filter(0.2 μm , Lot No. 610251CD, ADVANTEC, Japan)
- ⑤ Methanol(Lot No. 0000113893, J.T. Baker, USA)

(2) HPLC 분석조건

① Nitroglycerin

- Column: Sunfire C18(4.6 mm \times 150 mm, 3.5 μm , Waters, UK)
- Mobile phase: A(Acetonitrile 100 %)/B(Distilled water 100 %) = 55/45
- Flow rate: 1.0 mL/min
- Injection volume: 20 μL
- Column temperature: 40 °C
- Detector: PDA(205 nm)

② Cyclosporin A

- Column: Sunfire C18(4.6 mm \times 150 mm, 3.5 μm , Waters, UK)
- Mobile phase: A(Acetonitrile 100 %)/B(Distilled water 100 %) = 90/10
- Flow rate: 1.0 mL/min
- Injection volume: 10 μL

----- 다음페이지 계속 -----

시험성적서

성적서번호 : CT19-089635K

- Column temperature: 60 °C
- Detector: PDA(210 nm)

③ Tacrolimus

- Column: Sunfire C18(4.6 mm × 150 mm, 3.5 μm, Waters, UK)
- Mobile phase: A(Acetonitrile 100 %) = 100
- Flow rate: 1.0 mL/min
- Injection volume: 25 μL
- Column temperature: 40 °C
- Detector: PDA(210 nm)

3) 시험결과

(1) HPLC 분석법 검증(Table 1-1 ~ Table 1-3, Table 2-1 ~ Table 2-3, Figure 1 ~ Figure 3)

① 검량선의 직선성

분석법 검증 시 HPLC를 이용하여 시험물질 Nitroglycerin(1.000 μg/mL, 5.000 μg/mL, 10.000 μg/mL, 20.000 μg/mL, 100.000 μg/mL), Cyclosporin A(5.000 μg/mL, 10.000 μg/mL, 50.000 μg/mL, 100.000 μg/mL, 200.000 μg/mL, 1000.000 μg/mL) 및 Tacrolimus(2.500 μg/mL, 5.000 μg/mL, 10.000 μg/mL, 15.000 μg/mL, 20.000 μg/mL)의 검량선을 작성하였으며, R^2 는 각각 0.9999875, 0.9998628 및 0.9999240으로 직선성을 보였다.

② 정확도(Accuracy)

조제된 1.000 μg/mL 및 100.000 μg/mL 농도의 Nitroglycerin 표준용액을 각각 7 회 측정한 결과, 상대오차(% Error)는 각각 -8.4 % ~ -5.3 %, 0.1 % ~ 3.7 %의 결과를 보였으며, Cyclosporin A의 표준용액 1.000 μg/mL 및 100.000 μg/mL 농도를 각각 7 회 측정한 결과, 상대오차(% Error)는 각각 0.9 % ~ 4.2 %, -0.7 % ~ 2.7 %의 결과를 보였다. 조제된 1.000 μg/mL 및 100.000 μg/mL 농도의 Tacrolimus 표준용액을 각각 7 회 측정한 결과, 상대오차(% Error)는 각각 -3.8 % ~ 2.6 %, -3.0 % ~ 4.7 %의 결과를 보였다.

----- 다음페이지 계속 -----

시험성적서

성적서번호 : CT19-089635K

③ 정밀도(Precision)

조제된 1.000 µg/mL 및 100.000 µg/mL 농도의 Nitroglycerin 표준용액을 각각 7 회 측정한 결과, Nitroglycerin에 대한 변이계수(CV)는 각각 1.4 % 및 1.3 %의 결과를 보였으며, Cyclosporin A의 표준용액 1.000 µg/mL 및 100.000 µg/mL 농도를 각각 7 회 측정한 결과, Cyclosporin A에 대한 변이계수(CV)는 각각 1.3 % 및 1.2 %의 결과를 보였다. Tacrolimus에 대한 변이계수(CV)는 조제된 표준용액 1.000 µg/mL 및 100.000 µg/mL 농도를 각각 7 회 측정한 결과, 각각 2.5 % 및 2.3 %의 결과를 보였다.

④ 검출한계

Nitroglycerin, Cyclosporin A 및 Tacrolimus에 대한 정량분석 검증 결과, 검출한계는 각각 0.044 µg/mL, 0.212 µg/mL 및 0.204 µg/mL이었다.

⑤ 정량한계

Nitroglycerin, Cyclosporin A 및 Tacrolimus에 대한 정량분석 검증 결과, 정량한계는 각각 0.132 µg/mL, 0.642 µg/mL 및 0.618 µg/mL이었다.

(2) 시험결과(Table 3-1 ~ Table 3-3)

① Nitroglycerin

약물 Nitroglycerin의 수액 투브에 대한 흡착도 시험결과, 초기 조제한 수액의 약물을 측정농도(0 h) 및 수액 투브 통과 후 채취 시점(1 h, 2 h, 3 h, 4 h)에서의 약물을 측정농도는 각각 98.647 µg/mL, 95.740 µg/mL, 91.108 µg/mL, 93.360 µg/mL, 103.762 µg/mL의 농도분석 결과를 보였다. 각 시점 (1 h, 2 h, 3 h, 4 h) 별 개별 흡착도는 각각 2.9 %, 7.6 %, 5.4 %, 0 %의 결과를 보였다.

② Cyclosporin A

약물 Cyclosporin A의 수액 투브에 대한 흡착도 시험결과, 초기 조제한 수액의 약물을 측정농도(0 h) 및 수액 투브 통과 후 채취 시점(1 h, 2 h, 3 h, 4 h)에서의 약물을 측정농도는 각각 48.908 µg/mL, 46.647 µg/mL, 47.392 µg/mL, 48.026 µg/mL, 48.146 µg/mL의 농도분석 결과를 보였다. 각 시점 (1 h, 2 h, 3 h, 4 h) 별 개별 흡착도는 각각 4.6 %, 3.1 %, 1.8 %, 1.6 %의 결과를 보였다.

----- 다음페이지 계속 -----

시험성적서

성적서번호 : CT19-089635K

③ Tacrolimus

약물 Tacrolimus의 수액 투브에 대한 흡착도 시험결과, 초기 조제한 수액의 약물 측정농도(0 h) 및 수액 투브 통과 후 채취 시점(1 h, 2 h, 3 h, 4 h)에서의 약물 측정농도는 각각 9.581 µg/mL, 8.928 µg/mL, 9.083 µg/mL, 9.717 µg/mL, 9.693 µg/mL의 농도분석 결과를 보였다. 각 시점 (1 h, 2 h, 3 h, 4 h) 별 개별 흡착도는 각각 6.8 %, 5.2 %, 0 %, 0 %의 결과를 보였다.

4) 결론

약물 Nitroglycerin, Cyclosporin A 및 Tacrolimus에 대한 수액 투브의 약물 흡착도 시험 결과, 개별 시간(1 h, 2 h, 3 h, 4 h)에서의 평균 약물흡착도는 각각 4.0 %, 2.8 % 및 3.0 %의 결과를 보였다.

----- 다음페이지 계속 -----



시험성적서

성적서번호 : CT19-089635K

5) Tables

Table 1-1. Accuracy and precision of Nitroglycerin analysis using HPLC

Rep.	Concentration(μg/mL)		%Error ¹⁾ (%)	Mean % Error ²⁾	CV ³⁾ (%)	Mean CV (%)
	Nominal conc.	Measured conc.				
1		0.920	-8.0			
2		0.945	-5.5			
3		0.918	-8.2			
4	1.000	0.939	-6.1		1.4	
5		0.947	-5.3			
6		0.916	-8.4			
7		0.935	-6.5			
1		100.531	0.5	3.9		1.4
2		100.456	0.5			
3		100.371	0.4			
4	100.000	100.463	0.5		1.3	
5		100.103	0.1			
6		100.266	0.3			
7		103.748	3.7			

1) (Measured conc.-Nominal conc.)/Nominal conc. × 100

2) Mean of % Error absolute value

3) S.D. of measured conc./Mean of measured conc. × 100

----- 다음페이지 계속 -----



시험성적서

성적서번호 : CT19-089635K

Table 1-2. Accuracy and precision of Cyclosporin A analysis using HPLC

Rep.	Concentration(µg/mL)		%Error ¹⁾ (%)	Mean % Error ²⁾	CV ³⁾ (%)	Mean CV (%)
	Nominal conc.	Measured conc.				
1		5.131	2.6			
2		5.127	2.5			
3		5.149	3.0			
4	5.000	5.024	0.5		1.3	
5		5.045	0.9			
6		5.209	4.2			
7		5.154	3.1			
1		997.529	-0.2	1.6		1.3
2		997.180	-0.3			
3		992.997	-0.7			
4	1000.000	994.123	-0.6		1.2	
5		1027.249	2.7			
6		1005.149	0.5			
7		1005.244	0.5			

1) (Measured conc.-Nominal conc.)/Nominal conc. × 100

2) Mean of % Error absolute value

3) S.D. of measured conc./Mean of measured conc. × 100

—— 다음페이지 계속 ——



시험성적서

성적서번호 : CT19-089635K

Table 1-3. Accuracy and precision of Tacrolimus analysis using HPLC


Rep.	Concentration(µg/mL)		%Error ¹⁾ (%)	Mean % Error ²⁾ (%)	CV ³⁾ (%)	Mean CV (%)
	Nominal conc.	Measured conc.				
1		2.480	-0.8			
2		2.438	-2.5			
3		2.522	0.9			
4	2.500	2.565	2.6		2.5	
5		2.412	-3.5			
6		2.418	-3.3			
7		2.404	-3.8			
1		20.472	2.4	2.4		2.4
2.42		20.314	2.6			
3		20.229	1.1			
4	20.000	20.192	1.0		2.3	
5		19.398	-3.0			
6		20.439	2.2			
7		20.936	4.7			

1) (Measured conc.-Nominal conc.)/Nominal conc. × 100

2) Mean of % Error absolute value

3) S.D. of measured conc./Mean of measured conc. × 100

----- 다음페이지 계속 -----



시험성적서

성적서번호 : CT19-089635K

Table 2-1. LOD and LOQ of Nitroglycerin using HPLC

Rep.	Nominal conc. ($\mu\text{g/mL}$)	Area of Peak	S.D. ¹⁾	Slope of Calibration Curve ²⁾	R^2 of calibration curve	LOD ³⁾ ($\mu\text{g/mL}$)	LOQ ⁴⁾ ($\mu\text{g/mL}$)
1		38.772					
2		39.725					
3		38.676					
4	1.000	39.517	0.507	38.38296	0.9999875	0.044	0.132
5		39.793					
6		38.618					
7		39.326					

1) S.D. of peak area, 2) Slope of calibration curve for the analytical method validation

3) LOD: Limit of detection(Mean S.D./slope of calibration curve×3.3),

4) LOQ: Limit of quantitation(Mean S.D./slope of calibration curve×10)

Table 2-2. LOD and LOQ of Cyclosporin A using HPLC




Rep.	Nominal conc. ($\mu\text{g/mL}$)	Area of Peak	S.D. ¹⁾	Slope of Calibration Curve ²⁾	R^2 of calibration curve	LOD ³⁾ ($\mu\text{g/mL}$)	LOQ ⁴⁾ ($\mu\text{g/mL}$)
1		73.023					
2		72.968					
3		73.272					
4	5.000	71.500	0.916	14.26093	0.9998628	0.212	0.642
5		71.791					
6		74.127					
7		73.342					

1) S.D. of peak area, 2) Slope of calibration curve for the analytical method validation

3) LOD: Limit of detection(Mean S.D./slope of calibration curve×3.3)

4) LOQ: Limit of quantitation(Mean S.D./slope of calibration curve×10)

----- 다음페이지 계속 -----



시험성적서

성적서번호 : CT19-089635K

Table 2-3. LOD and LOQ of Tacrolimus using HPLC

Rep.	Nominal conc. ($\mu\text{g/mL}$)	Area of Peak	S.D. ¹⁾	Slope of Calibration Curve ²⁾	R^2 of calibration curve	LOD ³⁾ ($\mu\text{g/mL}$)	LOQ ⁴⁾ ($\mu\text{g/mL}$)
1		30.172					
2		29.541					
3		30.800					
4	2.500	31.458	0.934	15.11059	0.9999240	0.204	0.618
5		29.142					
6		29.227					
7		29.027					

1) S.D. of peak area, 2) Slope of calibration curve for the analytical method validation

3) LOD: Limit of detection(Mean S.D./slope of calibration curve $\times 3.3$)

4) LOQ: Limit of quantitation(Mean S.D./slope of calibration curve $\times 10$)

----- 다음페이지 계속 -----

시험성적서

성적서번호 : CT19-089635K

Table 3-1. Concentration analysis of Nitroglycerin in 5 % dextrose

Time (h)	Measured conc. ($\mu\text{g/mL}$)	Dilution effect	Compensated conc. ($\mu\text{g/mL}$) ¹⁾	Mean compensated conc. ($\mu\text{g/mL}$)	Absorption rate(%) ²⁾	Mean absorption rate(%)
0 h	19.744		98.720			
	19.676	5	98.380	98.647	-	
	19.768		98.840			
1 h	19.138		95.690			
	19.182	5	95.910	95.740	2.9	
	19.124		95.620			
2 h	18.404		92.020			
	18.113	5	90.565	91.108	7.6	4.0
	18.148		90.740			
3 h	19.331		96.655			
	18.350	5	91.750	93.360	5.4	
	18.335		91.675			
4 h	20.989		104.945			
	20.602	5	103.010	103.762	0 ³⁾	
	20.666		103.330			

1) Measured conc. ($\mu\text{g/mL}$) × Dilution effect

2) (Mean compensated conc. At 0 h - Mean compensated conc. at x h)/Mean compensated conc. at 0 h × 100

3) Negative value represented as 0

----- 다음페이지 계속 -----

시험성적서

성적서번호 : CT19-089635K

Table 3-2. Concentration analysis of Cyclosporin A in 5 % dextrose

Time (h)	Measured conc. ($\mu\text{g/mL}$)	Dilution effect	Compensated conc. ($\mu\text{g/mL}$) ¹⁾	Mean compensated conc. ($\mu\text{g/mL}$)	Absorption rate(%) ²⁾	Mean absorption rate(%)
0 h	24.452		48.904			
	24.747	2	49.494	48.908	-	
	24.163		48.326			
1 h	23.611		47.222			
	22.816	2	45.632	46.647	4.6	
	23.544		47.088			
2 h	24.198		48.396			
	23.034	2	46.068	47.392	3.1	2.8
	23.856		47.712			
3 h	24.725		49.450			
	24.408	2	48.816	48.026	1.8	
	22.906		45.812			
4 h	24.617		49.234			
	23.911	2	47.822	48.146	1.6	
	23.691		47.382			

1) Measured conc. ($\mu\text{g/mL}$) × Dilution effect

2) (Mean compensated conc. At 0 h - Mean compensated conc. at x h)/Mean compensated conc. at 0 h × 100

----- 다음페이지 계속 -----

시험성적서

성적서번호 : CT19-089635K

Table 3-3. Concentration analysis of Tacrolimus in 5 % dextrose

Time (h)	Measured conc. ($\mu\text{g/mL}$)	Dilution effect	Compensated conc. ($\mu\text{g/mL}$) ¹⁾	Mean compensated conc. ($\mu\text{g/mL}$)	Absorption rate(%) ²⁾	Mean absorption rate(%)
0 h	4.833	2	9.666	9.581	-	
	4.739		9.478			
	4.800		9.600			
1 h	4.401	2	8.802	8.928	6.8	
	4.429		8.858			
	4.562		9.124			
2 h	4.665	2	9.330	9.083	5.2	3.0
	4.459		8.918			
	4.501		9.002			
3 h	4.921	2	9.842	9.717	0 ³⁾	
	4.804		9.608			
	4.850		9.700			
4 h	4.815	2	9.630	9.693	0	
	4.834		9.668			
	4.891		9.782			

1) Measured conc. ($\mu\text{g/mL}$) × Dilution effect

2) Mean compensated conc. At 0 h - Mean compensated conc. at x h) / Mean compensated conc. at 0 h × 100

3) Negative value represented as 0

----- 다음페이지 계속 -----

시험성적서

성적서번호 : CT19-089635K

6) Figures

2019-06-Calibration.C:\Users\?????.Desktop\2019\CT19-CT19\CT19-089635_Nitroglycerin_Validation\CT19-089635_validation_20190814_NitroglycerinPage 1 of 1

Calibration Report

Nitroglycerin - Signal 1 = 4.053 nm.

Peak Type: Ordin
 Left Window: 0.2 min
 Right Window: 0.2 min
 Response Base: Area
 Curve Fit Type: Linear
 Origin: Zero Run Mean
 Weighting Method: None
 Subst. Equation: $Y = 38.38296 \times + 3.45618$
 Correlation Coef.: 0.999975
 Residuals: Y 10518 (mV.s)
 Uncertainton X: None
 Uncertainton Y: None



	Response	Amount	Resp. Factor	Per cent	Unit
1	37.4831	1.0000	0.0268	1	□
2	188.1003	5.0000	0.0268	1	□
3	367.8285	10.0000	0.0268	1	□
4	734.3688	20.0000	0.0268	1	□
5	3638.4435	100.0000	0.0260	1	□
6	0.0000	0.0000	2.0000	1	□
7	0.0000	0.0000	0.0000	0	□
8	0.0000	0.0000	0.0000	0	□
9	0.0000	0.0000	0.0000	0	□
10	0.0000	0.0000	0.0000	0	□
11	0.0000	0.0000	0.0000	0	□
12	0.0000	0.0000	0.0000	0	□
13	0.0000	0.0000	0.0000	0	□
14	0.0000	0.0000	0.0000	0	□
15	0.0000	0.0000	0.0000	0	□
16	0.0000	0.0000	0.0000	0	□
17	0.0000	0.0000	0.0000	0	□
18	0.0000	0.0000	0.0000	0	□
19	0.0000	0.0000	0.0000	0	□
20	0.0000	0.0000	0.0000	0	□
21	0.0000	Blank	0.0000	0	□



11/18-

Figure 1. Calibration report of Nitroglycerin standard solution in analysis method review

----- 다음페이지 계속 -----



시험성적서

성적서번호 : CT19-089635K


2019-0Calibration.C:\Users\777\Documents\2019\CT19-089635K\CT19-089635K_Cyclosporin A\Validation\CT19-089635_Validation_20190801_Cyclosporin Page 1 of 1

Calibration Report

Cyclosporin A - Signal 1 - 3.307 mm

Peak Type	Order
Left Window	0.2 mm
Right Window	0.2 mm
Response Base	Area
Curve Fit Type	Linear
UTRn	Zero not used
Weighting Method	None
Stsd. Equation	$Y = 14.26020^* X + 77.9801$
Correlation Coef.	0.9999928
Residual	83.09005 [mV s]
Linearization X	None
Linearization Y	None


	Retention Time	Amount	Resp Factor	Rec No	Used
1	72.39005	5.0000	0.0005	1	<input checked="" type="checkbox"/>
2	742.3238	10.0000	0.0703	1	<input checked="" type="checkbox"/>
3	771.7011	50.0000	0.0648	1	<input checked="" type="checkbox"/>
4	1552.3877	100.0000	0.0644	1	<input checked="" type="checkbox"/>
5	3030.3960	200.0000	0.0647	1	<input checked="" type="checkbox"/>
6	1431.4721	1000.0000	0.0659	1	<input checked="" type="checkbox"/>
7	d 0000	0.0000	0.0000	0	<input type="checkbox"/>
8	d 0000	0.0000	0.0000	0	<input type="checkbox"/>
9	d 0000	0.0000	0.0000	0	<input type="checkbox"/>
10	d 0000	0.0000	0.0000	0	<input type="checkbox"/>
11	d 0000	0.0000	0.0000	0	<input type="checkbox"/>
12	d 0000	0.0000	0.0000	0	<input type="checkbox"/>
13	d 0000	0.0000	0.0000	0	<input type="checkbox"/>
14	d 0000	0.0000	0.0000	0	<input type="checkbox"/>
15	d 0000	0.0000	0.0000	0	<input type="checkbox"/>
16	d 0000	0.0000	0.0000	0	<input type="checkbox"/>
17	d 0000	0.0000	0.0000	0	<input type="checkbox"/>
18	d 0000	0.0000	0.0000	0	<input type="checkbox"/>
19	d 0000	0.0000	0.0000	0	<input type="checkbox"/>
20	d 0000	0.0000	0.0000	0	<input type="checkbox"/>
21	d 0000	Blank	0.0000	0	<input type="checkbox"/>



2019

Figure 2. Calibration report of Cyclosporin A standard solution in analysis method review

----- 다음페이지 계속 -----



시험성적서

성적서번호 : CT19-089635K


2019-08-11Calibration C:\Users\777\OneDrive\Desktop\2019\CT19-G1\91\CT19-089635_Tacrolimus_validation\CT19-089635_validation_20190811_Tacrolimus.cdf Page 1 of 1

Calibration Report

Tacrolimus - Signal 1 - 2.7 min

Peak Type	<input checked="" type="checkbox"/> Order
Left Window	<input checked="" type="checkbox"/> 0.2 min
Right Window	<input checked="" type="checkbox"/> 0.2 min
Response Base	<input checked="" type="checkbox"/> Area
Curve Fit Type	<input checked="" type="checkbox"/> Linear
Units	<input checked="" type="checkbox"/> Zero not used
Weighting Method	<input checked="" type="checkbox"/> None
Stoel. Equation	$y = 15.11059 \times - 7.30569$
Correlation Coef	0.9999240
Medium	1.19277 [mV/s]
Unsaturation X	<input checked="" type="checkbox"/> None
Unsaturation Y	<input checked="" type="checkbox"/> None



	Peak/Order	Amount	Resp Factor	Run No.	Used
1	30.5687	2.5000	0.0818	1	<input checked="" type="checkbox"/>
2	68.3985	5.0000	0.0731	1	<input checked="" type="checkbox"/>
3	144.5715	10.0000	0.0892	1	<input checked="" type="checkbox"/>
4	217.1200	15.0000	0.0891	1	<input checked="" type="checkbox"/>
5	285.1480	20.0000	0.0875	1	<input checked="" type="checkbox"/>
6	0.0000	0.0000	0.0000	1	<input checked="" type="checkbox"/>
7	0.0000	0.0000	0.0000	0	<input type="checkbox"/>
8	0.0000	0.0000	0.0000	0	<input checked="" type="checkbox"/>
9	0.0000	0.0000	0.0000	0	<input checked="" type="checkbox"/>
10	0.0000	0.0000	0.0000	0	<input checked="" type="checkbox"/>
11	0.0000	0.0000	0.0000	0	<input checked="" type="checkbox"/>
12	0.0005	0.0000	0.0000	0	<input checked="" type="checkbox"/>
13	0.0000	0.0000	0.0000	0	<input checked="" type="checkbox"/>
14	0.0000	0.0000	0.0000	0	<input checked="" type="checkbox"/>
15	0.0000	0.0000	0.0000	0	<input checked="" type="checkbox"/>
16	0.0000	0.0000	0.0000	0	<input checked="" type="checkbox"/>
17	0.0000	0.0000	0.0000	0	<input checked="" type="checkbox"/>
18	0.0000	0.0000	0.0000	0	<input checked="" type="checkbox"/>
19	0.0000	0.0000	0.0000	0	<input checked="" type="checkbox"/>
20	0.0000	0.0000	0.0000	0	<input checked="" type="checkbox"/>
BL	0.0000	Blank	0.0000	0	<input checked="" type="checkbox"/>



10/14

Figure 3. Calibration report of Tacrolimus standard solution in analysis method review

----- 다음페이지 계속 -----



시험성적서

성적서번호 : CT19-089635K




Figure 4. Nitroglycerin standard solution(1.000 µg/mL) for analysis method review

-1/18-

----- 다음페이지 계속 -----




시험성적서

성적서번호 : CT19-089635K

2019-08-14 7:23 Chromatogram C:\Users\????\Desktop\AAAAA\Test\Calb\Nitroglycerin_validation_20190814_Nitroglycerin_STD_100 ug/mL.PRM Page 1 of 1

Analysis Report



6/14

Figure 5. Nitroglycerin standard solution(100.000 μg/mL) for analysis method review

----- 다음페이지 계속 -----



시험성적서

성적서번호 : CT19-089635K




Figure 6. Cyclosporin A standard solution(5.000 µg/mL) for analysis method review

----- 다음페이지 계속 -----



시험성적서

성적서번호 : CT19-089635K

2019-08-09 7 Chromatogram C:\ Users\ 7 7 7 1\Desktop\ AAAA\ Test\ Calb\ Cyclosporin A_validation_20190808_Cyclosporin A_STD_1000 ug_mL.PRM Page 1 of 1

Analysis Report

Chromatogram Info:


File Name: C:\ Users\ 7 7 7 1\Desktop\ AAAA\ Test\ Calb\ Cyclosporin A_validation_20190808_Cyclosporin A_STD_1000 ug_mL.PRM
 Origin: Acquired, Acquisition started 2019-08-08 7:8:35:36
 Project: C:\ Users\ 7 7 7 1\Desktop\ AAAA\ Projects\ Test.PJW

File Created: 2019-08-08 7:8:41:38
 Acquired Date: 2019-08-08 7:8:41:37
 By: Ethnol (Analyst)

Sample Info:

Sample ID: Cyclosporin A_validation_20190808
 Sample: Cyclosporin A_STD_1000 ug_mL
 Inj. Volume [mL]: 0.01

Amount [ug/mL]: 0
 I STD Amount: 0
 Dilution: 1




Result Table (Uncal - Calb, Cyclosporin A_validation_20190808_Cyclosporin A_STD_1000 ug_mL - Channel 1)

	Reten. Time [min]	Area [mV.s]	Height [mV]	Adu [%]	Height [%]	Wt/Wt [min]	Prec. Purity [-]	Compound Name
1	3.300	1444126	675705	100.0	100.0	0.38	99.9	
Total		1444126	675705	100.0	100.0			

49/48

Figure 7. Cyclosporin A standard solution(1000.000 µg/mL) for analysis method review

----- 다음페이지 계속 -----



시험성적서

성적서번호 : CT19-089635K




Figure 8. Tacrolimus standard solution(2.500 μg/mL) for analysis method review

88/128

----- 다음페이지 계속 -----



시험성적서

성적서번호 : CT19-089635K




Figure 9. Tacrolimus standard solution(20.000 µg/mL) for analysis method review


92/128

----- 다음페이지 계속 -----



시험성적서


성적서번호 : CT19-089635K



29/128


Figure 10. Nitroglycerin(100.000 µg/mL) in 5 % dextrose solution for sampling time at 4 h

----- 다음페이지 계속 -----



시험성적서


성적서번호 : CT19-089635K



83/14

Figure 11. Cyclosporin A(50.000 µg/mL) in 5 % dextrose solution for sampling time at 4 h

----- 다음페이지 계속 -----



시험성적서

성적서번호 : CT19-089635K





Figure 12. Tacrolimus(10.000 µg/mL) in 5 % dextrose solution for sampling time at 4 h

----- 다음페이지 계속 -----



시험성적서

성적서번호 : CT19-089635K




Figure 13. TPE tube

----- 끝 -----

